

File: No

Integration Events: AutoIntegrate

Generate Report During Run Method: No

Signal Correlation Window: 0.020

Quantitative Report Settings

Report Type: Summary

Output Destination

Screen: Yes
Printer: No
File: No

Generate Report During Run Method: No

Quantification T/E (3 points)

Calibration Last Updated: Mon Jul 24 17:15:46 2006

Reference Window: 2.00 Minutes
Non-Reference Window: 1.00 Minutes
Correlation Window: 0.10 minutes
Default Multiplier: 1.00
Default Sample Concentration: 0.00

Compound Information

-
1) 17méthyltestostérone (ISTD TR)
Ret. Time 20.94 min., Extract & Integrate from 20.44 to 21.44 min.

Lvl ID	Conc (ng/ml)	Response
1	100.000	5915926
2	100.000	4981326
3	100.000	5130998

ISTD conc: 100.000 ng/ml
Curve Fit: Linear

-
2) epitestosterone ()
Ret. Time 18.52 min., Extract & Integrate from 18.02 to 19.02 min.

Lvl ID	Conc (ng/ml)	Response
1	5.000	234336
2	30.000	1211449
3	60.000	2929592

Curve Fit: Linear, forced through origin
